

## Test Instruction

### Ovalbumin

96/48 Tests

Enzyme Immunoassay  
for the Quantitative  
Determination of  
Ovalbumin in Food

Cat.-No.: OVA-E01/E04

Version: February 21<sup>st</sup>, 2019

This document represents a combined test instruction for the products OVA-E01 (96 well) and OVA-E04 (48 well).

### Immunolab GmbH

Otto-Hahn-Str. 16  
D-34123 Kassel  
Germany

Phone: +49 (0) 561 491742-0  
Fax: +49 (0) 561 491742-20  
E-Mail: [info@immunolab.de](mailto:info@immunolab.de)

[www.immunolab.de](http://www.immunolab.de)

Sensitivity (Ovalbumin)	4 ppb
Recovery	95 - 115%
Incubation Time	60 min

### General Information

Hen's egg (*Gallus gallus*) is very rich of proteins and represents an important food source for humans. While proteins of egg yolk only have minor allergenicity, many proteins of egg white are known to be allergenic. In addition to lysozyme, ovomucoid, ovaltransferrin and livetin, ovalbumin represents the main fraction of the egg white allergens. Amongst others egg powder as well as pure ovalbumin is often used as fining reagent for wine. For allergic persons the consumption of ovalbumin represents a critical problem. Already very low amounts of the allergen can cause allergic reactions, which may lead to anaphylactic shock in severe cases. Because of this, ovalbumin allergic persons must strictly avoid the consumption of ovalbumin containing food. Non-declared addition of ovalbumin in food is hazardous for allergic people. Cross-contamination, mostly in consequence of the production process, is often noticed. Since July 1, 2012 the European Union requests allergen labeling for wine if milk or egg proteins are used during the production and are still present at a detection level of 0.25 mg/L or greater. Thus for the detection of ovalbumin residues, sensitive assay systems are required.

The **Immunolab Ovalbumin ELISA** represents a highly sensitive detection system for ovalbumin and is particularly capable of the quantification of ovalbumin residues in wine.

### Principle of the Test

The **Immunolab Ovalbumin** quantitative test is based on the principle of the enzyme linked immunosorbent assay. An antibody directed against ovalbumin is bound on the surface of a microtiter plate. Ovalbumin containing samples or standards are given into the wells of the microtiter plate. After 20 minutes incubation at room temperature, the wells are washed with diluted washing solution to remove unbound material. A peroxidase conjugated second antibody directed against ovalbumin is given into the wells, and after 20 minutes of incubation the plate is washed again. A substrate solution is added and incubated for 20 minutes, resulting in the development of a blue colour. The colour development is inhibited by the addition of a stop solution, and the colour turns yellow. The yellow colour is measured photometrically at 450 nm. The concentration of ovalbumin is directly proportional to the colour intensity of the test sample.

## Precautions

Full compliance of the following good laboratory practices (GLP) will determine the reliability of the results:

1. Prior to beginning the assay procedure, bring all reagents to room temperature (20-25°C).
2. All reagents should be mixed by gentle inversion or swirling prior to use. Do not induce foaming.
3. Once the assay has been started, all subsequent steps should be completed without interruption and within the recommended time limits.
4. Replace caps in all the reagents immediately after use. Do not interchange vial stoppers.
5. Use a separate disposable tip for each specimen to prevent cross-contamination.
6. All specimens and standards should be run at the same time, so that all conditions of testing are the same.
7. Do not mix components from different batches.
8. Do not use reagents after expiration date.
9. Check both precision and accuracy of the laboratory equipment used during the procedure (micropipets, ELISA reader etc.).

## Health and safety instructions

1. Do not smoke or eat or drink or pipet by mouth in the laboratory.
2. Wear disposable gloves whenever handling patient specimens.
3. Avoid contact of substrate and stop solution with skin and mucosa (possible irritation, burn or toxicity hazard). In case of contact, rinse the affected zone with plenty of water.
4. Handling and disposal of chemical products must be done according to good laboratory practices (GLP).

## Reagents

The kit contains reagents for 96/48 determinations. They have to be stored at 2-8°C. Expiry data are found on the labels of the bottles and the outer package.

1. Microtiter plate consisting of 12/6 strips with 8 breakable wells each, coated with anti-ovalbumin antibodies.
2. Ovalbumin standards (0; 25; 100; 250; 500 ppb of ovalbumin): 5 vials with 2.0 mL each, dyed red, ready-to-use
3. Conjugate (anti-ovalbumin-peroxidase): 15/7.5 mL, dyed red, ready-to-use.
4. Substrate Solution (TMB): 15 mL, ready-to-use.
5. Stop Solution (0.5 M H<sub>2</sub>SO<sub>4</sub>): 15 mL, ready-to-use.
6. Extraction and sample dilution buffer (Tris): 2/1 x 120 mL as 10x concentrate, dyed red. Dilute 1+9 with distilled water. Stored at 4°C the diluted buffer is stable for at least one week. If during the cold storage crystals precipitate, the concentrate should be warmed up to 37°C for 15 minutes.
7. Washing Solution (PBS + Tween 20): 60 mL as 10x concentrate. Dilute 1+9 with distilled water. Stored at 4°C the diluted buffer is stable for at least 4 weeks. If during the cold storage crystals precipitate, the concentrate should be warmed up to 37°C for 15 minutes.
8. Plastic bag to store unused microtiter strips.
9. Instruction Manual.

## Additional Instrumentation and Reagents (not provided)

### Instrumentation

- 100 - 1000 µL micropipets
- Volumetric flask
- Analytical balance
- Mortar, mixer
- Water bath
- Centrifuge
- ELISA reader (450 nm)

### Reagents

- double distilled water

## Sample Preparation

Due to high risk of cross-contamination all applied instruments like applicator, mortar, glass vials etc. have to be **cleaned thoroughly** before and after each sample. To identify possible cross-contamination caused by previous extractions it is strongly recommended to note the sequence of the extractions.

The following sample preparation should be applied for wine samples:

1. 1 ml of the wine sample is diluted in 19 mL of **pre-diluted** extraction buffer.
2. 100 µL of the wine solution are applied per well. If the results of a sample are out of the measuring range, further dilution with the **pre-diluted** extraction and sample dilution buffer is necessary. The additional dilution has to be considered when calculating the concentration.

The following sample preparation should be applied for all other kinds of samples:

1. To maximize homogeneity and representativeness of the sample drawing, a minimum of 5 g sample should be pulverized finely in a mortar, impact mill etc.
2. 1 g of the homogenized mixture is suspended in 20 mL of **pre-diluted** extraction buffer. Afterwards the suspension is incubated for 15 min in a pre-heated water bath at 60°C. To ensure good homogeneity, the samples should be shaken every two minutes.
3. The samples are centrifuged for 10 minutes at 2000 g. If it is not possible to separate the supernatant from the precipitate completely, the suspension should be filtrated if necessary.
4. 100 µL of particle-free solution are applied per well. If the results of a sample are out of the measuring range, further dilution with the **pre-diluted** extraction and sample dilution buffer is necessary. The additional dilution has to be considered when calculating the concentration.

## Procedure

The washing solution is supplied as 10x concentrate and has to be **diluted** 1+9 with double distilled water before use.

In any case the **ready-to-use** standards provided should be determined twofold. When samples in great quantities are determined, the standards should be pipetted once before the samples and once after the samples. For final interpretation the arithmetic mean is used for calculation.

In consideration of GLP and quality control requirements a duplicate measurement of samples is recommended.

The procedure is according to the following scheme:

1. Prepare samples as described above.
2. Pipet 100 µL **ready-to-use** standards or prepared samples in duplicate into the appropriate wells of the microtiter plate.
3. Incubate for 20 minutes at room temperature.
4. Wash the plate three times as follows: Discard the contents of the wells (dump or aspirate). Pipet 300 µL of diluted washing solution into each well. After the third repetition empty the wells again and remove residual liquid by striking the plate against a paper towel. The wash procedure is critical. Insufficient washing will result in poor precision and falsely elevated absorbencies.
5. Pipet 100 µL of conjugate (anti-ovalbumin-peroxidase) into each well.
6. Incubate for 20 minutes at room temperature.
7. Wash the plate as outlined in 4.
8. Pipet 100 µL of substrate solution into each well.
9. Allow the reaction to develop in the dark (e.g. cupboard or drawer; the chromogen is light-sensitive) for 20 minutes at room temperature.
10. Stop enzyme reaction by adding 100 µL of stop solution (0.5 M H<sub>2</sub>SO<sub>4</sub>) into each well. The blue colour will turn yellow upon addition.
11. After thorough mixing, measure absorbance at 450 nm (reference wavelength 620 nm), using an ELISA reader. The colour is stable for 30 minutes.

## Calculation of results

The ready-to-use standards are prepared for a direct determination of sample concentrations. The dilution of samples in the extraction process as described in the above stated sample preparation procedure is already considered. Additional dilution due to high sample concentration has to be accounted for.

1. Calculate the average optical density (OD 450 nm) for each set of reference standards or samples.
2. Construct a standard curve by plotting the mean optical density obtained for each reference standard against its concentration in ppb on semi-log graph paper with the optical density on the vertical (y) axis and the concentration on the horizontal (x) axis. Alternatively the evaluation can be carried out by software. In this case the 4-parameter method should be preferred.

- Using the mean optical density value for each sample, determine the corresponding concentration of ovalbumin in ppb from the standard curve. Depending on experience and/or the availability of computer capability, other methods of data reduction may be employed.

### Typical Standard Values

The following table contains an example for a typical standard curve. The binding is calculated as percent of the absorption of the 500 ppb standard. These values are only an example and should not be used instead of the standard curve which has to be measured in each new test.

Ovalbumin (ppb)	% binding of 500 ppb
500	100
250	55
100	26
25	9
0	1

### Performance

#### Sensitivity

The limit of detection (LOD) of the **Immunolab Ovalbumin test** is 4 ppb.

The limit of quantification (LOQ) of the **Immunolab Ovalbumin test** is 25 ppb.

Due to the variety of sample matrices and their influence on the blank, results less than the LOQ should be treated as negative.

#### Cross-reactivity

The Cross-reactivities of other eggwhite proteins as well as potential fining reagents was determined as follows:

Bovine Gelatin	0%
Conalbumin	< 0.2%
Egg White Protein (total)	75%
Fish	0%
Fish Gelatin	0%
Isinglass	0%
Lysozyme	< 0.02%
Non Fat Dry Milk	0%
Ovomucoid	< 0.02%

#### Precision

Intra-assay Precision	3 - 4%
Inter-assay Precision	2 - 3%

### Linearity

The serial dilution of spiked samples (white wine, red wine) resulted in a dilution linearity of 94% - 96%.

### Recovery

By spiking wine samples (white wine, rosé wine, red wine) with different amounts of ovalbumin, recoveries resulted in 95-115%.

### References

- Rolland JM, et al. (2008) - Specific and sensitive enzyme-linked immunosorbent assays for analysis of residual allergenic food proteins in commercial bottled wine. *J Agric Food Chem*, 56(2):349-54
- Kirschner S, et al. (2009) - Allergenicity of wine containing processing aids: A double-blind, placebo-controlled food challenge. *J Inv Clin Imm*, 19(3):210-17
- Weber P, et al. (2007) - Investigation of the allergenic potential of wines fined with various proteinogenic fining agents by ELISA- *J Agric Food Chem*, 55(8):3127-33
- Yeung JM, et al. (2000) - Determination of egg proteins in food products by enzyme immunoassay. *J AOAC Int*, 83(1):139-43
- Hefle SL, et al. (2001) - Development of a sandwich enzyme-linked immunosorbent assay for the detection of egg residues in processed foodstuff. *J Food Prot*, 64(11):1812-6
- Watanaba H, et al. (2005) - Study on detection of allergenic substances (egg and milk) in processed meat products and frozen foods. *Shokuhin Eiseigaku Zasshi*, 46(4):139-47
- Lifrani A, et al. (2009) - Development of animal models and sandwich-ELISA tests to detect the allergenicity and antigenicity of fining agent residues in wines. *J Agric Food Chem*, 57(2) :525-34